

HIOKI



ANALOG MΩ HiTESTER 3490

Field measuring instruments



3-range

INSULATION & CONTINUITY

3Ω range, 200mA



Bright LED



OPTION

TEST LEAD WITH
REMOTE CONTROL SWITCH 9788

Insulation Testing in 3 Easy Steps

Flip the Cover, Select Range & Test



ISO 9001
JMI-0216



ISO14001
JQA-E-90091



www.hioki.com

HIOKI company overview, new products, environmental considerations and other information are available on our website.



Luminous Scale

Check for Live Circuits



The LIVE CIRCUIT LED will light up in red whenever the voltage exceeds 20V AC between the LINE and EARTH terminals, and when at least 20V DC is still remaining during the auto discharge.

Check the Battery Status



BATT : Hi **BATT : Lo** **BATT : Dead**

Be well-informed about the condition of your batteries. Green signals that the battery level is sufficiently high, and red warns of low battery power. Replace the batteries before the LED turns completely off.

General Specifications

Power source	Rated power voltage: 1.5 VDC × 4, LR6 alkaline battery × 4
Continuous operating time	Approx. 20 hours (at 500V range, no load)
Auto Power Save	Approx. 15 minutes
Operating Temperature & Humidity	0 to 40°C (32 to 104°F), 90%RH or lower (non-condensating) 40 to 50°C (104 to 122°F), at 50°C and below relative with linear decrease up to 50%RH
Storage Temperature & Humidity	0 to 50°C (32 to 122°F), 90%RH or lower (non condensating)
Maximum rated voltage to earth	600 VAC, Measurement Category III, Anticipated Transient Overvoltage: 6000V
Dielectric strength	5312 VAC, 50/60Hz, Measurement terminals - electrical enclosure, current sensitivity 1 mA
Standards	Safety EN61010 EMC EN61326
Drop Proof	On concrete: 1m/1 time
Dimensions (excluding protrusions)	Approx. 159W × 162H × 48D mm (6.26"W × 6.38"H × 1.89"D)
Mass	Approx. 550g (19.4 oz.) (including battery, not including test lead)
Accessories	9787 Test Lead, Instruction manual, Shoulder strap, LR6 alkaline battery × 4

3490 specifications Guaranteed for one year at 23°C±5°C (73°F±9°F) and 90% RH

	Rated output voltage		
	250 V DC	500 V DC	1000 V DC
Insulation Resistance Measurement	Effective maximum indicated value	100 MΩ	
	Center scale value	1 MΩ	
	1st effective measuring range	0.05 to 50 MΩ	
		±5% of indicated value	
	2nd effective measuring range	0.01 to 0.05 MΩ	
		50 to 100 MΩ	
	±10% of indicated value		
	Open circuit voltage	1 to 1.2 times of rated output voltage	
Lower limit measurement resistance value to be maintained reted output voltage	0.25 MΩ	0.5 MΩ	1 MΩ
Rated current	1mA (Tolerance: 1 to 1.2 times of the rating value)		
Overload protection	1200V AC (10 sec.)		
Resistance Measurement	Ranges	3Ω	30Ω
	Center scale value	1.5Ω	15Ω
	Accuracy	±0.09Ω	
	Open-circuit voltage	4.1 to 6.9 V	
	Measuring current	200mA DC or more	20mA DC or more
	Overload protection	720V AC (10 sec., by Fuse)	
AC voltage Measurement	Measuring range	0 to 600 V (50/60 Hz)	
	Accuracy	±5% of maximum scale value	
	Input resistance	100 kΩ or more (50/60Hz)	
	Overload protection	1200V AC (10 sec.)	

Accessories

TEST LEAD 9787 (1.2m)



Options

TEST LEAD WITH REMOTE CONTROL SWITCH 9788 (1.2m)




TEST PROBE 9294 (1.2m)



CONNECTION CORD 9257 (1.3m)



BREAKER PIN 9288 (for Model 9294)



HIOKI

HIOKI E. E. CORPORATION

HEAD OFFICE :

81 Koizumi, Ueda, Nagano, 386-1192, Japan
TEL +81-268-28-0562 / FAX +81-268-28-0568
E-mail: os-com@hioki.co.jp

HIOKI USA CORPORATION :

6 Corporate Drive, Cranbury, NJ 08512 USA
TEL +1-609-409-9109 / FAX +1-609-409-9108
E-mail: hioki@hiokiusa.com

HIOKI (Shanghai) Sales & Trading Co., Ltd. :

1904 Shanghai Times Square Office, 93 Huai Hai Zhong Road
Shanghai, P.R.China POSTCODE: 200021
TEL +86-21-6391-0090/0092 FAX +86-21-6391-0360
E-mail: info-sh@hioki.cn

Beijing Office :

A-2602 Freetown, 58 Dong San Huan Nan Road
Beijing, P.R.China POSTCODE: 100022
TEL +86-10-5867-4080/4081 FAX +86-10-5867-4090
E-mail: info-bj@hioki.cn

Guangzhou Office :

Room 303, Profit Plaza, No.76, West Huangpu Road
Guangzhou, P.R.China POSTCODE: 510623
TEL +86-20-38392673/2676 FAX +86-20-38392679
E-mail: info-gz@hioki.cn

DISTRIBUTED BY